




झिमरुक गाउँपालिका
JHIMRUK RURAL MUNICIPALITY



गाउँ कार्यपालिकाको कार्यालय

OFFICE OF THE RURAL MUNICIPAL EXECUTIVE



लुम्बिनी प्रदेश, नेपाल
LUMBINI PROVINCE, NEPAL

आ.ब: २०८३/०८४

च.नं. ३८६

मिति: २०८३/०६/२३

विषय: दररेट उपलब्ध गराउने सम्बन्धी सूचना ।

उपरोक्त सम्बन्धमा आ.व. २०८३/०८४ मा यस झिमरुक गाउँपालिका अन्तर्गत रहेका स्वास्थ्य सस्थाहरूका लागि Digital X-RAY Machine आवश्यक भएकोले उल्लेखित स्पेशलिफिकेन बमोजिमको सामग्रीहरू आफुले दिन सकिने दररेटमा सात (७) दिन भित्र पेश गर्नुहुन सम्बन्धित आपूर्तिकर्ता, फर्म र सरोकारवालाहरूका लागि यो सूचना प्रकाशित गरिएको छ । दररेट तपसिलको मेल आई. डि. मा पठाउन सकिने छ ।

तपसिल

S.N.	Name of Equipment	Description	Unit price Per Unit
1.	DIGITAL X-RAY WITH 100MA MACHINE	<u>Digital Radiography (DR) System with Printer</u> Digital radiography (DR) is an advanced form of x-ray inspection which produces a digital radiographic image instantly on a computer, Flat Panel digital detector with work station and dry film printer. It shall be suitable to be used for adult and paediatric patients in general Radiography examination, Detector type: electronic flat CsI detector-wireless and Indirect Conversion Method, Portable wireless and wired of min. 35x43cm size detector, Pixel size must be 150 micron or less, Battery type: Lithium Ion, or better, All-in-One software integrates not only X-ray image exposure and image acquisition but also mini-PACS functions such as viewing and filing, main console PC shall have capacity to store 1TB of images, have image filter to enhance image quality, Printer: Double Tray Online Dry Thermal/laser Printer, Performance (max.) 14x17inch: Approx. 50 sheets/hour, Film size 14x17inch, 10x12inch, 8x10 inch etc., Day light Film Loading, one-year warranty, Having ISO and CE for Flat Panel and printer and USFDA for DR Console software and panel. <u>Portable X-ray Machine, 100mA</u> A general purpose mobile X-ray machine 100 mA HF or more with Horizontal Bucky Table, suitable to be used for adult and paediatric patient in general radiography examination and it shall operate on single AC power supply, Microprocessor-controlled high frequency generator at least 40KHz, Power Output: not less than 6.0kW or bet-	

“झिमरुक गाउँपालिकाको समृद्धिको आधार: व्यवसायिक कृषि, पर्यटन तथा वाणिज्य, सुरक्षित आवास र विकासको सी.”

इमेल : jhimrukmun@gmail.com , info@jhimrukmun.gov.np वेबसाइट: jhimrukmun.gov.np

निमित्त प्रमुख प्रशासकीय अधिकारी



		ter, Control panel with total soft feather touch operations, Anatomical Programmable Radiographic(APR) mode shall be available, Digital Display of KV and mAs, Anode Heat Storage Capacity min.: 30KJ, unit shall have counter balanced arm system/spring balance arm system, supply horizontal bulky table (HBT), wall mounted chest stand, Unit having 1 years warranty, having ISO13485, EU_CE certificate or USFDA approved product certificate, X-ray system should be BIS and AERB Approved or equivalent.	
2.	DIGITAL X-RAY WITH 500MA MACHINE	<u>Digital Radiography (DR) System with Printer</u> Digital radiography (DR) is an advanced form of x-ray inspection which produces a digital radiographic image instantly on a computer, Flat Panel digital detector with work station and dry film printer. It shall be suitable to be used for adult and paediatric patients in general Radiography examination, Detector type: electronic flat CsI detector-wireless and Indirect Conversion Method, Portable wireless and wired of min. 35x43cm size detector, Pixel size must be 150 micron or less, Battery type: Lithium Ion, or better, All-in-One software integrates not only X-ray image exposure and image acquisition but also mini-PACS functions such as viewing and filing, main console PC shall have capacity to store 1TB of images, have image filter to enhance image quality, Printer: Double Tray Online Dry Thermal/laser Printer, Performance (max.) 14x17inch: Approx. 50 sheets/hour, Film size 14x17inch, 10x12inch, 8x10 inch etc., Day light Film Loading, one-year warranty, Having ISO and CE for Flat Panel and printer and USFDA for DR Console software and panel. <u>X-ray Machine 500 mA</u> General purpose X-ray machine 500 mA, digital controlled high frequency generator, output 40 KW or above to give a constant output suitable for radiography, KV range: 40KV to 125KV at 500mA, Control: Digital, Must Display mA, mAs & kVP, Rotating anode X-ray tube with dual focal spot, Collimator with auto/manual shut off facility, should have over voltage /current protection, soft touch key selection, kV, mA & mAs, Should have anatomical programmed radiography, manual/APR user selection, should have floor mounted ceiling free tube stand and ceiling free fusion DR system. HBT should have glass top that ensure proper centering and also for	



झिमरुक गाउँपालिका
JHIMRUK RURAL MUNICIPALITY

गाउँ कार्यपालिकाको कार्यालय
OFFICE OF THE RURAL MUNICIPAL EXECUTIVE

भ्यागुते, प्युठान
BHYAGUTE, PYUTHAN

लुम्बिनी प्रदेश, नेपाल
LUMBINI PROVINCE, NEPAL

		DR FPD solution, Standard accessories like Lead apron lightweight with 2mm thickness- 01no., one-year warranty, Having ISO, CE, BIS/AERB or USFDA approved Certificate.	
			TOTAL
			VAT
			GRAND TOTAL

Note : kindly mentioned if any deviation

email : info@jhimrukmun.gov.np

: jhimrukmun@gmail.com

विकास के.सी.
निमित्त प्रमुख प्रशासकीय अधिकृत

“झिमरुक गाउँपालिकाको समृद्धिको आधार: व्यवसायिक कृषि, पर्यटन तथा वाणिज्य, सुरक्षित आवास र पूर्वाधार”

इमेल : jhimrukmun@gmail.com , info@jhimrukmun.gov.np वेबसाइट: jhimrukmun.gov.np